


MELLES GRIOT



Filter I.D.

Test Instrument:
Cary 5E (0905)

Test Technician:
P. Melo /cc

Date:

Customer Name:

P.O. Number:
136777

Pert Number:
03 FIM 006

Serial Number:
L036-07

Lot Number:
21697

Filter Data

Center Wavelength:
547.64 nm

Bandwidth:

50%:	9.62 nm
10%:	14.18 nm
1%:	21.15 nm

Peak Transmission:
59.48%

FWHM Coordinates:

542.828 nm
552.449 nm

Angle of Incidence:
0.00°

Ambient Temperature:
23.0° C.